

Written by S. Hudan (shudan@indiana.edu) and Z. Gosser (zgosser@indiana.edu), 8/06/2009

<u>Detector ID</u>	<u>dE/E</u>	<u>Thickness (<math>\mu\text{m}</math>)</u>	<u>Micron Depletion (V)</u>	<u>IU Depletion (V)</u>	<u>IU Current (<math>\mu\text{A}</math>)</u>
1661-10	dE	65	25	25	0.42
1652-04	dE	64	19	19	0.7
1652-06	dE	69	19	-	-
1652-07	dE	67	19	19	0.31
1161-04	dE	68	19	19	0.71
1652-05	dE	67	19	19	0.48
1661-02	dE	67	21	21	0.98
1661-01	dE	66	18	18	0.19
1661-09	dE	70	19	-	-
1661-05	dE	66	-	-	-
1652-01	dE	66	-	-	-
1890-03	dE	68	-	-	-
0862-16	dE	305	-	-	-
1747-06	dE	65	-	19	0.93
1653-13	E	497	-	-	-
1194-17	E	471	30	66	1.31
1742-07	E	510	45	-	-
1742-04	E	499	45	55	0.92
1133-20	E	461	56	66	7.49
1140-19	E	481	55	65	3.25
1140-20	E	478	-	-	-
1133-17	E	463	-	-	-
1133-21	E	477	-	-	-
1140-25	E	463	-	-	-
1140-18	E	456	50	60	5.29
1140-21	E	458	75	85	1.52
1140-17	E	476	48	-	-
1133-18	E	482	-	-	-
1133-19	E	482	-	-	-
1673-05	E	978	-	-	-
1674-09	E	480	-	50	0.7
1674-12	E	482	-	60	0.96

Notes:

The IU depletion is the voltage used in E432 (GANIL 07)

The IU current is the current at the end of E432 for the given voltage